

EDF-9999

Method 1613 Calibration Solutions
[CS1-CS5]5 X 0.2 mL
in Nonane

All concentrations are in ng/mL

Unlabeled	CS0.1*	CS0.2*	CS0.5*	CS1	CS2	CS3	CS4	CS5
2,3,7,8-Tetrachlorodibenzo- <i>p</i> -dioxin	0.05	0.1	0.25	0.5	2.0	10	40	200
2,3,7,8-Tetrachlorodibenzofuran	0.05	0.1	0.25	0.5	2.0	10	40	200
1,2,3,7,8-Pentachlorodibenzo- <i>p</i> -dioxin	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,7,8-Pentachlorodibenzofuran	0.25	0.5	1.25	2.5	10	50	200	1000
2,3,4,7,8-Pentachlorodibenzofuran	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,4,7,8-Hexachlorodibenzo- <i>p</i> -dioxin	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,6,7,8-Hexachlorodibenzo- <i>p</i> -dioxin	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,7,8,9-Hexachlorodibenzo- <i>p</i> -dioxin	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,4,7,8-Hexachlorodibenzofuran	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,6,7,8-Hexachlorodibenzofuran	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,7,8,9-Hexachlorodibenzofuran	0.25	0.5	1.25	2.5	10	50	200	1000
2,3,4,6,7,8-Hexachlorodibenzofuran	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,4,6,7,8-Heptachlorodibenzo- <i>p</i> -dioxin	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,4,6,7,8-Heptachlorodibenzofuran	0.25	0.5	1.25	2.5	10	50	200	1000
1,2,3,4,7,8,9-Heptachlorodibenzofuran	0.25	0.5	1.25	2.5	10	50	200	1000
Octachlorodibenzo- <i>p</i> -dioxin	0.5	1.0	2.50	5.0	20	100	400	2000
Octachlorodibenzofuran	0.5	1.0	2.50	5.0	20	100	400	2000
¹³ C-Labeled	CS0.1*	CS0.2*	CS0.5*	CS1	CS2	CS3	CS4	CS5
1,2,3,4-Tetrachlorodibenzo- <i>p</i> -dioxin (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
2,3,7,8-Tetrachlorodibenzo- <i>p</i> -dioxin (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
2,3,7,8-Tetrachlorodibenzo- <i>p</i> -dioxin (³⁷ Cl ₄ , 96%)	0.05	0.1	0.25	0.5	2.0	10	40	200
2,3,7,8-Tetrachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,7,8-Pentachlorodibenzo- <i>p</i> -dioxin (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,7,8-Pentachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
2,3,4,7,8-Pentachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,4,7,8-Hexachlorodibenzo- <i>p</i> -dioxin (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,6,7,8-Hexachlorodibenzo- <i>p</i> -dioxin (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,7,8,9-Hexachlorodibenzo- <i>p</i> -dioxin (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,4,7,8-Hexachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,6,7,8-Hexachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,7,8,9-Hexachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
2,3,4,6,7,8-Hexachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,4,6,7,8-Heptachlorodibenzo- <i>p</i> -dioxin (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,4,6,7,8-Heptachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
1,2,3,4,7,8,9-Heptachlorodibenzofuran (¹³ C ₁₂ , 99%)	100	100	100	100	100	100	100	100
Octachlorodibenzo- <i>p</i> -dioxin (¹³ C ₁₂ , 99%)	200	200	200	200	200	200	200	200

*EDF-9999 Calibration Series contains levels CS1-CS5. Levels CS0.1, CS0.2, and CS0.5 are available separately.